

ASTM D4501 TESTING FIXTURE

SHEAR STRENGTH OF ADHESIVE BONDS BETWEEN RIGID SUBSTRATES BY THE BLOCK-SHEAR METHOD



ASTM D4501 Testing Fixture - Application

This test method describes a procedure and fixture used to determine shear strengths of adhesives used to bond materials with moduli higher than the modulus of the adhesive. The size and shape of the specimens are variable within the physical restraints of the fixture.

Two blocks of specific sizes are bonded together. A typical test specimen consists of a 1" x 1" block, typically up to ½" thick, bonded in the center of a 3" x 3" block, also typically up to ½" thick. The block thicknesses are not critical. To maintain the largest block in position in the main body, against the front plate, is used the toggle clamp. The smaller block is positioned in the opening in the guillotine bar, which is temporarily held in the proper position for specimen insertion by the thumb screw shown in the front face plate.

Before the test is started, this thumb screw is loosened. The fixture is gripped at bottom and top in standard tensile grips in the testing machine.

Additional Information

Test Standard	ASTM D4501
Maximum Load	45 kN
Temperature Range	-80 °C to 130 °C
Specimen Thickness	depending on material
Mass	10 Kg

References:

ASTM Standard D 4501-01 (2001), "Standard Test Method for Shear Strength of Adhesive Bonds Between Rigid Substrates by the Block-Shear Method, American Society for Testing and Materials, West Conshohocken, Pennsylvania (first published in 1985).

☎ Sòphia High Tech S.r.l.

Legal Office: Viale Manzoni 113 - int.2 80040 - Poggiomarino (NA) - Italy
Headquarter: Via Romani 228, 80048 - Sant'Anastasia (NA) - Italy
Tel: +39 0823504748 **Fax:** +39 0818651210
mail: info@sophiahightech.com **P.IVA/VAT:** 07547751219
www.sophiahightech.com

☎ Sòphia CZ S.r.o.

Headquarter / Legal Office: Jičínská 226/17, Praha 3 130 00 - Czech Republic
Tel: +420 773462127
mail: info@sophiahightech.com
IC: 05617979
www.sophiahightech.com